

DICOM Correction Proposal Form

Correction Number		CP-139 (Letter Ballot)
Log Summary: Specify Photometric Interpretation for CR IOD		
Type of Modification	Name of Standard	
Addition		
Rationale for Correction		
<p>The CR IOD does not restrict Photometric Interpretation, but is obviously only applicable to grayscale images. This CP strictly speaking changes the meaning of the CR SOP Class but has no effect in practice.</p> <p>The type of the attribute is already Type 1 in the General Image Module.</p>		
Sections of documents affected		
C.8.1.2 CR Image Module - Table C.8-2		
Correction Wording:		
Append to C.8.1.2 CR Image Module - Table C.8-2 CR Image Module Attributes:		

C.8.1.2 CR Image Module

...

**Table C.8-2
CR IMAGE MODULE ATTRIBUTES**

Attribute Name	Tag	Type	Attribute Description
Photometric Interpretation	(0028,0004)	1	<p>Specifies the intended interpretation of the pixel data. Shall have one of the following Enumerated Values:</p> <p>MONOCHROME1</p> <p>MONOCHROME2</p>